



OFFICE OF THE PRINCIPAL  
DIAMOND HARBOUR GOVERNMENT MEDICAL COLLEGE  
DEPARTMENT OF HEALTH AND FAMILY WELFARE  
GOVERNMENT OF WEST BENGAL

Ph: 03174-255-633

E-mail: diaharmc@gmail.com

MEMO No. DHGMC/2021/882

Date: 06/07//2021

**Corrigendum Notice No:1**


(Ref: NIT No: 16/2021-2022, e-tender No: DHGMC/2021/871 dt; 02/07/2021)

In reference to the above said matter, the below mentioned corrigendum are hereby included in the bid document (DHGMC/2021/871 dt; 02/07/2021).

As per unanimous decision of the tender and purchase committee meeting 06/07/2021 in presence of the all members of the Tender and Purchase committee DHGMCH, Diamond Harbour and TIA, DHGMCH, Diamond Harbour.

• **Addendum Specification of HD Endoscopy set with Monitor, Light Source, Camera & Recording device:**

1. The above mentioned equipment's technical specification has been uploaded wrongly which was unintentionally. The fresh addendum specification is annexed (Annexure I) with this notice.
2. Last date for submission of the above mentioned e-tender 19/07/2021 insted of 10/07/2021 and technical bid opening date 21/07/2021 instead of 12/07/2021.
3. For Details and Downloading the tender documents, parties or agencies interested in participation in E-Tender process may please visit website: <https://wbtenders.gov.in> (organization- Health & Family Welfare Department). For any further assistance, please visit office of the Principal, Diamond Harbour Govt. Medical College & Hospital, Diamond Harbour, South 24 Parganas.PIN-743331.

  
Principal


Diamond Harbour Govt. Medical College.  
Diamond Harbour, South 24 Pgs.

Memo No: DHGMC/2021/ 882 /1/(8)

Dated:- 06/07/2021

Copy forwarded for information and necessary action to the :-

1. The Director of Medical Education & Ex-Officio Secretary to the Govt. of West Bengal, Dept. of Health and Family Welfare, Swasthya Bhawan, 6N-29, Sector-V, Salt Lake City, Kolkata-700 091.
2. The District Magistrate, South 24 pgs .
3. The Medical Superintendent-cum-Vice Principal, DHGMCH, South 24 pgs
6. The Sub Divisional Officer , Diamond Harbor, South 24 Parganas
7. The Chairman, Diamond Harbour Municipality, South 24 Parganas
5. The Accounts Officer., DHGMC, S 24 Pgs
6. All members of the Tender & Purchase Committee, DHGMCH, S 24Pgs
7. IT Cell, Department of Health & F.W. Swasthya Bhawan- with a request to upload this notice in the Health Departmental website.
8. The District Information Officer, NIC, South 24 pgs Collectorate Bldg., Alipore -with a request upload this notice in the Departmental website.

  
Principal

Diamond Harbour Govt. Medical College.  
Diamond Harbour, South 24 Pgs.

Principal  
Diamond Harbour Government Medical College & Hospital  
Diamond Harbour, South 24 Parganas





**TECHNICAL SPECIFICATION OF SURGICAL INSTRUMENTS FOR MUCOR  
DEBRIDEMENT SURGERY**

**A. TECHNICAL SPECIFICATION OF DESKTOP COMPOSITE NASAL ENDOSCOPY SYSTEM**

- The system should be compact and portable suitable for a variety of endoscopic applications from physician's offices to operating rooms in a variety of specialties.
- The powerful all-in-one unit should consist of everything needed for endoscopic imaging, Video recording, the monitor, camera, and light source.
- IT should have USB ports and a SD card slot for documentation purpose.
- It should also have the provision for Stroboscope for future upgradation.

S. No	Specification														
1	<b>High definition Camera System</b>														
	<p>The system should have the maximum Resolution of 1920 X 1080 pixels, progressive scan and the consistent use of 16: 9 formats</p> <ul style="list-style-type: none"><li>• Internal Medical grade recorder for Capturing FULL HD Videos and HD Stills with storage capability of 50GB internally and 500 GB externally.</li><li>• 18 inch or more touch screen for Play back of Recorded Videos and Controlling the Features camera system.</li><li>• System should have facility of controlling the light source and recording device from the camera head.</li><li>• Video Scopes can be connected to the same.</li></ul> <p><b><u>Technical Specifications:</u></b></p> <table><tr><td>Image sensor:</td><td>CMOS</td></tr><tr><td>Pixels</td><td>1920 x 1080</td></tr><tr><td>AGC:</td><td>Microprocessor controller</td></tr><tr><td>Control buttons:</td><td>3 (2 of them freely programmable)</td></tr><tr><td>Input: generator.</td><td>Keyboard input for character</td></tr><tr><td>Power Supply:-</td><td>100-240 VAC 50/60 Hz</td></tr><tr><td>Certified to :</td><td>IEC 601-1, 601-2-18, CSA 22.2 No. 601, UL 2601 and CE according to MDD, protection class1/CF</td></tr></table>	Image sensor:	CMOS	Pixels	1920 x 1080	AGC:	Microprocessor controller	Control buttons:	3 (2 of them freely programmable)	Input: generator.	Keyboard input for character	Power Supply:-	100-240 VAC 50/60 Hz	Certified to :	IEC 601-1, 601-2-18, CSA 22.2 No. 601, UL 2601 and CE according to MDD, protection class1/CF
Image sensor:	CMOS														
Pixels	1920 x 1080														
AGC:	Microprocessor controller														
Control buttons:	3 (2 of them freely programmable)														
Input: generator.	Keyboard input for character														
Power Supply:-	100-240 VAC 50/60 Hz														
Certified to :	IEC 601-1, 601-2-18, CSA 22.2 No. 601, UL 2601 and CE according to MDD, protection class1/CF														

**PROF. & HOD**  
DEPT.....  
Diamond Harbour Govt. Medical College & Hospital  
South 24 Parganas

A.Ghosh-02/7/21  
For, PROF. & HOD  
DEPT.....F.M.T.....  
Diamond Harbour Govt. Medical College & Hospital  
South 24 Parganas





2	<b>Inbuilt High Definition Medical Grade Monitor</b>
	The monitor should have: HDTV display in original 16: 9 format. 1080 p/ 50 & 1080 p/60 displays possible. LED crystal display. Max. Resolution of 1920X1080. Screen diagonal – 18 inches or more Integrated monitor
3.	<b>Inbuilt LED Light Source with Fiber optic cable</b>
	<ul style="list-style-type: none"><li>• Color Temperatures 6400K</li><li>• Integrated light source</li><li>• Light Outlets – 1</li><li>• Light Intensity Adjustment :- Continuously adjustable manually and automatically by cameras video output signal.</li><li>• Should be supplied with Diameter 4.8mm, Length 300cm.</li><li>• Certified To :- IEC 601-1 &amp; UL 544 CE According to MDD , protection class 1/CF</li></ul>
4. (a)	<b>Scope of supply should include of :</b> 1:Complete Integrated Desktop Endoscopy Terminal with accessories 2:One USB of minimum 8GB capacity 3:Fiber Optic Light cable Dia 3.5-4mm, length 230-250 cms 4:All above equipments should be from the same manufacturer with USFDA and European CE Certification.

A. Ghosh  
02/7/2021  
FOY, PROF. & HOD  
DEPT.....F.N.I.....  
Diamond Harbour Govt. Medical College & Hospital  
South 24 Parganas

**PROF. & HOD**  
DEPT.....  
Diamond Harbour Govt. Medical College & Hospital  
South 24 Parganas